# BASICS TO ADVANCED ANALYTICAL CHEMISTRY

*Yasmeen JUNEJO*
*Mehmet ÖZASLAN*

## Principles of Solvent Extraction

Solvent extraction is a method of isolating and purifying compounds or elements from a mixture. It involves the transfer of a solute from one solvent phase to another solvent phase, typically from an aqueous phase to an organic phase or vice versa. Solvent extraction is based on the principles of distribution or partitioning, where the solute distributes itself between two immiscible solvents, depending on its solubility and the solvents' properties.

The solvent extraction process consists of several steps, including mixing the two phases, allowing them to separate, and finally collecting and recovering the desired component from the solvent phase. The partition coefficient, Kd, defined as the ratio of the solute concentration in the organic phase to that in the aqueous phase at equilibrium, describes the solute's distribution between the two phases.

Solvent extraction is commonly used to purify metals from their ores, such as the extraction of copper from copper ores using an organic solvent. The process involves leaching the ore with an aqueous solution containing the metal ion, followed by solvent extraction of the metal ion using an organic solvent such as an alkyl phosphine oxide.

The selection of the solvent system is crucial in the solvent extraction process. The solvent should be immiscible with the initial solvent phase and have a high affinity for the solute to ensure efficient extraction. The solvent should also be easy to separate from the solute and not react with the solute or other components of the mixture.

Solvent extraction is also used in the pharmaceutical industry to purify natural products and isolate drug compounds from impurities. It is a useful technique for useful basic, and neutral compounds from complex mixtures, such as plant extracts.

In conclusion, solvent extraction is a widely used separation technique based on the principles of distribution or partitioning. It is a useful technique for purifying metals from their ores, isolating drug compounds from impurities, and separating compounds from complex mixtures. The solvent system selection is a critical factor affecting the efficiency and selectivity of the solvent extraction process.

## Analytical Separations

Analytical separations are techniques used to separate and identify components of a mixture based on their physical and chemical properties. Analytical separations are used in various fields, including chemistry, biology, and environmental science, to determine the composition of complex mixtures.

There are several types of analytical separations, including chromatography, electrophoresis, and mass spectrometry. Each technique involves different principles and methods of separation and detection.

Chromatography is a separation technique based on the differential partitioning of a sample between two phases, one stationary and one mobile. The sample is introduced onto the stationary phase, which may be a solid or liquid and is then separated by the mobile phase, which is typically a liquid or a gas. The separated components of the sample are detected as they emerge from the column or other separation device. Chromatography can be further classified into different types, including gas chromatography (GC), liquid chromatography (LC), and ion chromatography (IC).

Electrophoresis is a technique that separates charged particles, such as proteins or DNA, in an electric field. The sample is placed on a gel matrix, and an electric field is applied, causing the charged particles to migrate through the gel at different rates. The separated components of the sample can be detected using various methods, including staining or labeling.

Mass spectrometry is a technique that separates and detects ions based on their mass-to-charge ratio. The sample is ionized, and the resulting ions are separated in a magnetic field based on their mass-to-charge ratio. The separated ions are then detected and quantified.

Analytical separations have many applications in industry, research, and medicine. For example, chromatography is used to produce pharmaceuticals to separate and purify drug compounds. Electrophoresis is used in genetics and genomics research to separate and analyze DNA fragments. Mass spectrometry is used in proteomics and metabolomics to identify and quantify proteins and small molecules in complex mixtures.

In conclusion, analytical separations are a vital set of techniques used to separate and identify components of complex mixtures based on their physical and chemical properties. These techniques are widely used in various fields, including chemistry, biology, and environmental science, to determine the composition of complex mixtures. The selection of the appropriate technique for a particular application depends on the nature of the sample and the information required.

## Multiple Batch Extraction

Batch extraction is the process of simultaneously extracting multiple pieces of data from a source. It is a common task in data analysis and is often used to extract data from databases, web pages, or other sources. Multiple batch extraction involves extracting data from multiple sources or multiple parts of a source in one go.

1. **Identify the sources:** The first step in multiple batch extraction is identifying the sources from which you want to extract data. This could be multiple databases, web pages, or different sections of a single web page.
2. **Define the extraction criteria:** Once you have identified the sources, you must define the criteria for extracting data. This could include specifying the fields or columns you want to extract, setting filters or search criteria, or defining the range of data you want to extract.
3. **Choose a tool:** Many tools are available for batch extraction, ranging from simple web scrapers to more complex data extraction software. Choose a tool that fits your needs and is compatible with the sources you are working with.
4. **Set up the extraction process:** Depending on the tool you are using; you may need to set up the extraction process by configuring the tool to connect to the source and specifying the extraction criteria.

5. **Run the extraction:** Once the extraction process is set up, you can run the extraction and collect the data. Depending on the size of the data set, this may take some time.
6. **Clean and analyze the data:** After the data has been extracted, you may need to clean it up and organize it for analysis. This could involve removing duplicates, formatting the data, or merging it with other data sets. Once the data is clean, you can analyze it to gain insights or create reports.

Overall, multiple batch extraction can be a powerful tool for data analysis, allowing you to extract and analyze large amounts of data quickly and efficiently. However, it requires careful planning and attention to detail to ensure the data is extracted accurately and useful for analysis.

## Counter Current Distribution

Counter current distribution is a separation technique used in organic chemistry to separate components of a mixture. It involves the repeated extraction of a mixture with two immiscible solvents, each of which is placed in a separate phase. The process is based on mass transfer and partitioning principles, allowing for the isolation of individual components in a mixture.

1. **Principles of mass transfer and partitioning:** Counter-current distribution is based on the principles of mass transfer and partitioning. When a mixture is in contact with two immiscible solvents, its components partition the two phases based on their solubility in each solvent. This partitioning process is governed by the solvents' relative concentrations and the components' distribution coefficients in each solvent.
2. **The process:** The counter-current distribution process involves the repeated extraction of a mixture with two immiscible solvents, each of which is placed in a separate phase. The process is typically carried out in a series of separatory funnels containing the mixture and the two solvents. The two solvents are added to the funnel, which is shaken to mix the two phases. The mixture is then allowed to settle, separating the two phases.
3. **Repeated extraction:** The separated phases are then transferred to the next funnel in the series and repeated. The solvents are added to the funnel

containing the separated phases from the previous funnel, and the funnel is shaken to mix the phases. The mixture is then allowed to settle, and the phases are separated again. This process is repeated until the desired separation is achieved.

4. **Advantages of counter-current distribution**: Counter-current distribution has several advantages over other separation techniques. It is a simple and inexpensive process that does not require specialized equipment. It is also a highly efficient technique, allowing for separating individual components in a mixture with a high degree of resolution.
5. **Limitations of counter-current distribution:** While counter-current distribution is a highly effective technique, it does have some limitations. The process is time-consuming, as it requires repeated extractions. Its ability to separate components with similar partition coefficients is also limited.

Overall, counter-current distribution is a useful technique for separating components of a mixture based on their partitioning behaviour. While it does have some limitations, it is a powerful tool in the hands of an experienced organic chemist.

## Solvent Extraction by Flow Injection Method

Solvent extraction is a separation technique to isolate and purify compounds from complex mixtures. It involves using a solvent to selectively extract the target compound while leaving other components of the mixture behind. The flow injection method is a commonly used technique for performing solvent extraction, which involves using an automated system to deliver the sample to the solvent rapidly.

1. **The principle of solvent extraction:** Solvent extraction is based on the principle of partitioning, where the target compound selectively dissolves into the solvent while other components of the mixture are insoluble. This is achieved by mixing the sample with a suitable solvent and allowing the two phases to separate. The target compound selectively dissolves into the solvent, which can then be separated from the other components of the mixture.
2. **The flow injection method:** The flow injection method involves using an automated system to rapidly deliver the sample to the solvent. The sample is injected into a sample loop, which is then automatically switched to the solvent. The sample is then mixed with the solvent, allowing the target

compound to dissolve into the solvent selectively. The two phases are then separated, and the target compound can be further purified if necessary.

3. **Advantages of solvent extraction by flow injection method:** Solvent extraction by flow injection method has several advantages over other solvent extraction techniques. It is a rapid and efficient technique that can be automated, allowing for high-throughput analysis of samples. It also allows for the selective extraction of target compounds, resulting in higher purity and better recovery.
4. **Applications of solvent extraction by flow injection method:** Solvent extraction by flow injection method is commonly used in the analysis of environmental samples, such as water and soil, as well as in the analysis of pharmaceuticals and natural products. It is also used to analyse food and beverage samples to detect contaminants and additives.
5. **Limitations of solvent extraction by flow injection method:** Solvent extraction by flow injection method does have some limitations. The choice of solvent can greatly impact the efficiency and selectivity of the extraction. Additionally, the high flow rates used in the technique can lead to poor contact between the sample and the solvent, resulting in poor recovery and selectivity.

Overall, solvent extraction by flow injection method is a powerful technique for isolating and purifying target compounds from complex mixtures. Its ability to automate the process and deliver high-throughput results makes it a valuable tool in many areas of analytical chemistry.

## Solid-Phase Extraction by Flow Injection Method

Solid-phase extraction (SPE) is a technique to isolate and purify compounds from complex mixtures. It involves the uusinglid-phase material, such as a resin or silica gel, to selectively retain the target compound while other components of the mixture are washed away. The flow injection method is a commonly used technique for performing SPE, which involves using an automated system to rapidly deliver the sample to the solid-phase material.

1. **The principle of SPE**: Solid-phase extraction is based on the principle of partitioning, where the target compound selectively adsorbs onto the solid-

phase material. At the same time, other components of the mixture are washed away. This is achieved by passing the sample through a column containing the solid-phase material, allowing the target compound to interact with the stationary phase selectively.

2. **The flow injection method:** The flow injection method involves using an automated system to rapidly deliver the sample to the solid-phase material. The sample is injected into a sample loop, which is then automatically switched to the SPE column. The sample is then passed through the column, allowing the target compound to adsorb onto the solid-phase material selectively. The unbound components are washed away, and the target compound is eluted from the column using a suitable solvent.
3. **Advantages of SPE by flow injection method:** SPE by flow injection method has several advantages over other SPE techniques. It is a rapid and efficient technique that can be automated, allowing for high-throughput analysis of samples. It also allows for the selective extraction of target compounds, resulting in higher purity and better recovery.
4. **Applications of SPE by flow injection method:** SPE by flow injection method is commonly used in the analysis of environmental samples, such as water and soil, as well as in the analysis of pharmaceuticals and natural products. It also analyses food and beverage samples to detect contaminants and additives.
5. **Limitations of SPE by flow injection method:** SPE by flow injection method does have some limitations. The choice of solid-phase material and solvent can greatly impact the efficiency and selectivity of the extraction. Additionally, the high flow rates used in the technique can lead to channelling and poor contact between the sample and the solid-phase material, resulting in poor recovery and selectivity.

Overall, SPE by flow injection method is a powerful technique for isolating and purifying target compounds from complex mixtures. Its ability to automate the process and deliver high-throughput results makes it a valuable tool in many areas of analytical chemistry.

## Principles of Chromatography

1. Chromatography is a separation process frequently employed in analytical chemistry to separate and purify complicated mixtures. Chromatography's fundamental premise is based on the differential distribution of components in a combination of mobile and stationary phases. The stationary phase is often a solid or liquid coated on a solid support, while the mobile phase is a fluid that travels through it. The mixture's components interact with the mobile and stationary phases differentially, resulting in unequal migration through the stationary phase.
2. **Types of chromatography**: Chromatography is classified into three types: gas chromatography (GC), liquid chromatography (LC), and ion chromatography (IC). The chromatography method used is determined by the nature of the sample being analysed as well as the qualities of the target substances.
3. **Mobile phase:** The mobile phase is the fluid that flows through the stationary phase, carrying the mixture's components with it. The mobile phase used is determined by the nature of the sample being analysed as well as the qualities of the target molecules. The mobile phase in LC is often a solvent or a mixture of solvents, whereas the mobile phase in GC is typically an inert gas.
4. **Stationary phase:** The stationary phase is the solid or liquid phase that is coated on a solid support and interacts with the mixture's components. The type of the material being analysed, as well as the properties of the target chemicals, influence the choice of stationary phase. The stationary phase in LC is often a solid or liquid coated on a solid support, whereas the stationary phase in GC is typically a high boiling liquid coated on a solid support.
5. **Retention time:** The retention time is the time that a component of the mixture spends in the stationary phase before the mobile phase elutes it. The retention time is a function of the properties of the component, the properties of the stationary and mobile phases, and the flow rate of the mobile phase.
6. **Separation factor:** The separation factor is the degree of separation between two components in a mixture. It is a function of the retention times of the two components and is a measure of the selectivity of the chromatographic system.
7. **Detection:** Chromatography is typically coupled with a detection method, such as mass spectrometry, UV-Vis spectroscopy, or fluorescence spectroscopy, to identify and quantify the components of the mixture.

Overall, chromatography is a powerful separation technique that is widely used in analytical chemistry. The chromatography principles involve the differential distribution of components between the mobile and stationary phases, which depends on the properties of the components and the stationary and mobile phases.

## Structure of Chromatography

Chromatography is a separation technique used in various scientific disciplines to separate and analyze complex mixtures of compounds. The technique relies on the differential interactions between a sample and a stationary phase (solid or liquid) and a mobile phase (gas or liquid), separating the sample into its individual components. Chromatography can be divided into several categories based on the nature of the stationary phase, the separation mechanism, and the mode of operation.

The basic structure of chromatography consists of the following components:

1. **Stationary phase:** The stationary phase is immobilised solid or liquid material on a support, such as a column or a thin layer on a solid substrate. The stationary phase can be selected based on its capacity to interact selectively with the sample components.
2. The mobile phase is the gas or liquid that transports the sample through the stationary phase. The mobile phase can be chosen based on its capacity to dissolve sample components as well as its interactions with the stationary phase.
3. **Sample:** The sample is the mixture of compounds being analyzed by chromatography. The sample can be injected into the chromatography system using a syringe or an automatic sampler.
4. **Detector:** The detector is the device that measures the concentration of the sample components as they exit the chromatography column. The detector can be chosen based on the properties of the sample components, such as their absorbance, fluorescence, or mass-to-charge ratio.
5. **Data analysis:** The data generated by the detector can be analyzed using various methods, such as peak integration, spectral deconvolution, and statistical analysis. The data analysis can provide information about the sample components' identity, purity, and quantity.

The structure of chromatography varies greatly depending on the type of chromatography. The stationary phase in liquid chromatography, for example, can be a solid particle or a liquid deposited onto a solid support, whereas the stationary phase in gas chromatography is often a thin film sprayed onto a solid support. Additionally, the mode of operation can vary, such as in column chromatography, where the sample is separated by passing through a column packed with the stationary phase, or in thin-layer chromatography, where the sample is separated by migrating through a thin layer of stationary phase coated onto a solid substrate.

Overall, the structure of chromatography is diverse and can be adapted to suit a wide range of applications, making chromatography a powerful and versatile technique for the separation and analysis of complex mixtures of compounds.

## Classification of Chromatographic Techniques

Chromatography is a powerful separation process that is frequently utilised in a variety of scientific areas such as chemistry, biochemistry, and biotechnology. Chromatographic techniques rely on the differential interactions between a sample and a stationary phase (solid or liquid), separating the sample into its components.

Chromatographic techniques are grouped into numerous types based on the nature of the stationary phase, the mechanism of separation, and the mode of operation. In general, chromatography is divided into two types: liquid chromatography and gas chromatography.

1. **Liquid chromatography (LC)**

Liquid chromatography is a separation technique that uses a liquid stationary phase to separate a sample into its constituents. Liquid chromatography is divided into the following categories:

- **Normal phase chromatography:** The stationary phase is polar (e.g., silica gel) while the mobile phase is nonpolar (e.g., hexane) in normal phase chromatography. This chromatography method is effective for separating polar substances.
- **Reverse phase chromatography:** The stationary phase in reverse phase chromatography is nonpolar (e.g., C18) while the mobile phase is polar (e.g.,

water). This chromatography method is excellent for separating nonpolar substances.

- **Ion exchange chromatography:** In ion exchange chromatography, the stationary phase includes charged groups (for example, DEAE cellulose) while the mobile phase contains ions with opposing charges. This chromatography method is effective for separating charged compounds like proteins and nucleic acids.
- **Size exclusion chromatography:** In size exclusion chromatography, the stationary phase is porous, and the mobile phase is a buffer solution. This type of chromatography separates molecules based on their size.
- **Affinity chromatography:** In affinity chromatography, the stationary phase contains a ligand that specifically binds to the molecule of interest (e.g., an antibody). This type of chromatography is helpful for purifying proteins and other biomolecules.

2. **Gas chromatography (GC)**

Gas chromatography is a separation technique that involves using a gas as the mobile phase and a solid or liquid stationary phase to separate the sample into its individual components. Gas chromatography can be further classified into the following categories:

- **Gas-solid chromatography:** In gas-solid chromatography, the stationary phase is a solid (e.g., silica gel) and the mobile phase is a gas (e.g., helium). This type of chromatography is useful for separating volatile organic compounds.
- **Gas-liquid chromatography:** In gas-liquid chromatography, the stationary phase is a liquid (e.g., a silicone oil) coated onto a solid support (e.g., diatomaceous earth) and the mobile phase is a gas (e.g., helium). This type of chromatography is useful for separating nonpolar and moderately polar compounds.

Overall, chromatography is a versatile technique that can be used for a wide range of applications, including drug discovery, environmental analysis, and protein purification.

## Thin layer chromatography

Thin layer chromatography (TLC) is a chromatographic technique for separating and analysing compound mixtures. TLC is a rapid, easy, and low-cost technique that is widely used in chemistry, biochemistry, and environmental science.

TLC operates on the same basic concept as other chromatographic techniques in that a sample is separated based on its interactions with a stationary and mobile phase. A thin layer of a solid material, such as silica gel, alumina, or cellulose, is deposited onto a flat surface in TLC, such as a glass plate or a plastic sheet. The mobile phase is a solvent that moves up the thin layer due to capillary action.

The steps involved in performing TLC are as follows:

1. **Prepare the TLC plate:** A thin layer of the stationary phase is coated onto a flat surface, such as a glass plate or a plastic sheet. The thickness of the layer can vary depending on the type of TLC being performed, but it is typically between 0.1 and 0.5 mm.
2. **Prepare the sample:** The sample is dissolved in a solvent that is compatible with the mobile phase and spotted onto the TLC plate using a micropipette or a capillary tube.
3. **Develop the TLC plate:** The TLC plate is placed in a container with a little amount of mobile phase that is allowed to travel up the plate via capillary action. The mobile phase interacts with the stationary phase as it advances up the plate, causing the components of the sample to segregate depending on their differential interactions with the stationary phase.
4. **Visualize the separation:** The TLC plate is taken from the container and dried once the mobile phase has reached the top. By exposing the plate to a visualisation reagent, such as iodine, ninhydrin, or UV light, the separated components can be seen. Each component will appear as a spot on the plate, with its position and intensity depending on its interactions with the stationary phase.

It is used as a preliminary separation step before more advanced techniques, such as HPLC or GC. TLC is relatively easy to perform and requires minimal equipment, making it an ideal technique for routine analysis and screening of large samples.

## Column Chromatography

Column chromatography is a common laboratory technique used to separate and purify chemical compounds from a complex mixture. The technique relies on the differential interactions between the sample components and the stationary phase, which is packed into a long, narrow column. The separated components are collected as they elute from the column and can be further analyzed or purified.

The basic steps involved in performing column chromatography are as follows:

1. **Choose the stationary phase:** The stationary phase is often a solid material packed into the column, such as silica gel, alumina, or a polymer. The stationary phase chosen is determined by the type of sample being separated and the desired separation method.
2. **Pack the column:** The stationary phase is packed into a long, narrow column using a slurry of the stationary phase in a suitable solvent. The column is typically packed from the bottom to the top, with a frit or other filter at the bottom to prevent the stationary phase from escaping.
3. **Load the sample:** The sample is dissolved in a suitable solvent and loaded onto the top of the column. The sample is then allowed to enter the column and interact with the stationary phase.
4. **Elute the column:** The sample is eluted from the column by passing a solvent through the column. Depending on the desired separation mechanism, the solvent can be either a single solvent or a gradient of solvents. As the solvent passes through the column, the sample components interact with the stationary phase and are separated based on their differential interactions.
5. **Collect the fractions:** The separated components are collected as they elute from the column using a fraction collector or manually collecting the fractions. The fractions can be further analyzed or purified as needed.

Column chromatography can be performed using different modes of operation, such as gravity flow, flash chromatography, or preparative chromatography. Gravity flow chromatography is the simplest mode but is slow and requires careful flow rate control. Flash chromatography is a faster mode that uses compressed air to force the solvent through the column. Preparative chromatography is a larger-scale version of column chromatography used to purify larger quantities of material.

Overall, column chromatography is a powerful and versatile technique widely used in various scientific disciplines for separating and purifying chemical compounds.

## Ion Exchange Chromatography

Ion exchange chromatography (IEC) is a method of separating ions depending on their charge. The approach is often used to purify and separate biomolecules with a net charge at a certain pH, such as proteins and nucleic acids. The principle of IEC is based on the reversible binding of ions to charged stationary phases, which can be either cationic or anionic.

The basic steps involved in performing IEC are as follows:

1. **Choose the stationary phase:** The stationary phase is typically a resin that is charged either positively (cation exchange) or negatively (anion exchange). The choice of stationary phase depends on the type of ions being separated and the desired separation mechanism.
2. **Pack the column:** The stationary phase is packed into a column, similar to the process of column chromatography. The sample is then loaded onto the top of the column.
3. **Wash the column:** The column is washed with a buffer solution that matches the pH and ionic strength of the sample. This step ensures the stationary phase is fully equilibrated and ready to bind the sample ions.
4. **Elute the column:** The sample is eluted from the column by passing a salt or pH gradient. As the salt or pH changes, the ions in the sample are displaced from the stationary phase and eluted from the column. The eluted fractions can be collected and further analyzed or purified.

IEC can be performed in different modes of operation, such as analytical, preparative, and continuous ion exchange chromatography. Analytical IEC is used to separate and analyse small quantities of ions or biomolecules. In contrast, preparative IEC is used to purify and isolate larger quantities of biomolecules. Continuous ion exchange chromatography is used for the production of large-scale biopharmaceuticals.

IEC has several advantages, including high resolution and specificity, compatibility with a wide range of biomolecules, and the ability to handle complex mixtures.

However, the technique has limitations, such as low sample loading capacity and the potential for non-specific binding. IEC is a powerful and widely used technique for separating and purifying charged biomolecules.

## Electrophoresis

Electrophoresis is a technique used to separate charged molecules based on their size and charge. It is commonly used in molecular biology and biochemistry to separate and analyze nucleic acids, proteins, and other charged molecules. The principle of electrophoresis is based on the migration of charged particles in an electric field, where the charge, size, and shape of the molecules determines the direction and rate of migration.

The basic steps involved in performing electrophoresis are as follows:

1. **Choose the electrophoresis medium:** The electrophoresis medium is a gel or a solution that contains a matrix of charged particles, such as agarose or polyacrylamide. The choice of medium depends on the type of molecules being separated and the desired resolution.
2. **Prepare the sample:** The sample is prepared by mixing the charged molecules with a buffer solution and a loading buffer that contains a tracking dye. The loading buffer helps to visualize the sample during electrophoresis.
3. **Load the sample:** The sample is loaded onto the electrophoresis medium using a pipette or a gel loading apparatus. The sample is loaded into a well or a slot in the medium.
4. **Apply the electric field:** The electrophoresis medium is placed in a chamber filled with buffer solution and connected to a power supply. When the electric field is applied, the charged molecules in the sample migrate through the medium towards the oppositely charged electrode.
5. **Visualize the separated molecules:** After electrophoresis, the separated molecules can be visualized using a staining solution, such as ethidium bromide for nucleic acids or Coomassie blue for proteins. The separated molecules can also be transferred to a membrane for further analysis.

Electrophoresis can be performed using different modes of operation, such as **agarose gel electrophoresis, polyacrylamide gel electrophoresis (PAGE), and capillary electrophoresis**. Agarose gel electrophoresis is used to separate large nucleic acids, such as genomic DNA, while PAGE is used to separate proteins and smaller nucleic acids. Capillary electrophoresis is a high-resolution technique used to separate small molecules, such as amino acids and carbohydrates.

Electrophoresis has several advantages, including high resolution, sensitivity, and versatility. However, the technique also has limitations, such as the potential for band distortion, variable band intensity, and the need for specialized equipment. Overall, electrophoresis is a powerful and widely used technique for the separation and analysis of charged molecules.

## Multiple Batch Extraction

Batch extraction is the process of simultaneously extracting multiple pieces of data from a source. It is a common task in data analysis and is often used to extract data from databases, web pages, or other sources. Multiple batch extraction involves extracting data from multiple sources or multiple parts of a source in one go.

7. **Identify the sources:** The first step in multiple batch extraction is identifying the sources from which you want to extract data. This could be multiple databases, web pages, or different sections of a single web page.
8. **Define the extraction criteria:** Once you have identified the sources, you must define the criteria for extracting data. This could include specifying the fields or columns you want to extract, setting filters or search criteria, or defining the range of data you want to extract.
9. **Choose a tool:** Many tools are available for batch extraction, ranging from simple web scrapers to more complex data extraction software. Choose a tool that fits your needs and is compatible with the sources you are working with.
10. **Set up the extraction process:** Depending on the tool you are using, you may need to set up the extraction process by configuring the tool to connect to the source and specifying the extraction criteria.

11. **Run the extraction:** Once the extraction process is set up, you can run the extraction and collect the data. Depending on the size of the data set, this may take some time.
12. **Clean and analyze the data:** After the data has been extracted, you may need to clean it up and organize it for analysis. This could involve removing duplicates, formatting the data, or merging it with other data sets. Once the data is clean, you can analyze it to gain insights or create reports.

Overall, multiple batch extraction can be a powerful tool for data analysis, allowing you to extract and analyze large amounts of data quickly and efficiently. However, it requires careful planning and attention to detail to ensure that the data is extracted accurately and is useful for analysis.

## Analytical Separations

Analytical separations are techniques used to separate and identify components of a mixture based on their physical and chemical properties. Analytical separations are used in a variety of fields, including chemistry, biology, and environmental science, to determine the composition of complex mixtures.

There are several types of analytical separations, including chromatography, electrophoresis, and mass spectrometry. Each technique involves different principles and methods of separation and detection.

Chromatography is a separation technique based on the differential partitioning of a sample between two phases, one stationary and one mobile. The sample is introduced onto the stationary phase, which may be a solid or a liquid, and is then separated by the mobile phase, which is typically a liquid or a gas. The separated components of the sample are detected as they emerge from the column or other separation device. Chromatography can be further classified into different types, including gas chromatography (GC), liquid chromatography (LC), and ion chromatography (IC).

Electrophoresis is a technique that separates charged particles, such as proteins or DNA, in an electric field. The sample is placed on a gel matrix, and an electric field is applied, causing the charged particles to migrate through the gel at different rates.

The separated components of the sample can be detected using a variety of methods, including staining or labeling.

Mass spectrometry is a technique that separates and detects ions based on their mass-to-charge ratio. The sample is ionized, and the resulting ions are separated in a magnetic field based on their mass-to-charge ratio. The separated ions are then detected and quantified.

Analytical separations have many applications in industry, research, and medicine. For example, chromatography is used in the production of pharmaceuticals to separate and purify drug compounds. Electrophoresis is used in genetics and genomics research to separate and analyze DNA fragments. Mass spectrometry is used in proteomics and metabolomics to identify and quantify proteins and small molecules in complex mixtures.

In conclusion, analytical separations are a vital set of techniques used to separate and identify components of complex mixtures based on their physical and chemical properties. These techniques are widely used in various fields, including chemistry, biology, and environmental science, to determine the composition of complex mixtures. The selection of the appropriate technique for a particular application depends on the nature of the sample and the information required.

## Principles of solvent extraction

Solvent extraction is a method of isolating and purifying compounds or elements from a mixture. It involves the transfer of a solute from one solvent phase to another solvent phase, typically from an aqueous phase to an organic phase or vice versa. Solvent extraction is based on the principles of distribution or partitioning, where the solute distributes itself between two immiscible solvents, depending on its solubility and the solvents' properties.

The solvent extraction process consists of several steps, including mixing the two phases, allowing them to separate, and finally collecting and recovering the desired component from the solvent phase. The partition coefficient, Kd, which is defined as the ratio of the concentration of the solute in the organic phase to that in the aqueous phase at equilibrium, describes the solute's distribution between the two phases.

Solvent extraction is commonly used in the purification of metals from their ores, such as the extraction of copper from copper ores using an organic solvent. The process involves leaching the ore with an aqueous solution containing the metal ion, followed by solvent extraction of the metal ion using an organic solvent such as an alkyl phosphine oxide.

The selection of the solvent system is crucial in the solvent extraction process. The solvent should be immiscible with the initial solvent phase and have a high affinity for the solute to ensure efficient extraction. The solvent should also be easy to separate from the solute and not react with the solute or other components of the mixture.

Solvent extraction is also used in the pharmaceutical industry for the purification of natural products and isolation of drug compounds from impurities. It is a useful technique for the separation of acidic, basic, and neutral compounds from complex mixtures, such as plant extracts.

In conclusion, solvent extraction is a widely used separation technique based on the principles of distribution or partitioning. It is a useful technique for the purification of metals from their ores, the isolation of drug compounds from impurities, and the separation of compounds from complex mixtures. The selection of the solvent system is a critical factor that affects the efficiency and selectivity of the solvent extraction process.

**Analytical spectrophotometry** is a technique used in analytical chemistry to determine the concentration and properties of a substance in a sample by measuring the absorption or transmission of electromagnetic radiation. This technique involves using a spectrophotometer, which measures light intensity as a function of wavelength or frequency.

There are several types of analytical spectrophotometry, each of which is suited to measuring specific properties of a sample. Some of the most common types of spectrophotometry include:

1. **Ultraviolet-visible (UV-Vis) spectroscopy:** measures the absorption of ultraviolet and visible light by a sample, and is used to determine the concentration of chromophores, such as proteins and nucleic acids.
2. **Infrared (IR) spectroscopy:** measures the absorption of infrared radiation by a sample, and is used to identify functional groups in organic molecules and to determine the structure of complex molecules, such as proteins and polymers.
3. **Fluorescence spectroscopy:** measures the emission of light by a sample after it has been excited by a light source, and is used to study the properties of fluorescent molecules, such as fluorescent dyes used in biological imaging.
4. **Atomic absorption spectroscopy (AAS):** measures the absorption of light by free atoms in a sample, and is used to determine the concentration of metals in a sample.

Analytical spectrophotometry is a widely used and powerful technique in analytical chemistry, allowing for the determination of the concentration and properties of a wide range of analytes in various types of samples. It has numerous applications in fields such as environmental analysis, clinical chemistry, and biochemistry.

## Examples of Analytical Spectrophotometry

Here are some examples of analytical spectrophotometry:

1. **Protein quantification using UV-Vis spectroscopy:** UV-Vis spectroscopy is commonly used to quantify the concentration of proteins in a sample. This technique is based on the fact that proteins contain aromatic amino acids (such as tyrosine and tryptophan) that absorb light in the UV-Vis range. By measuring the absorbance of light at a specific wavelength (e.g., 280 nm), the concentration of protein in a sample can be determined.
2. **Identification of unknown organic compounds using IR spectroscopy:** IR spectroscopy is a powerful tool for identifying unknown organic compounds. Each functional group in an organic compound absorbs infrared radiation at a specific frequency, which produces a unique fingerprint that can be used for identification. The compound can be identified by comparing the IR spectrum of an unknown compound to a database of known spectra.

3. **Quantification of DNA using UV-Vis spectroscopy:** UV-Vis spectroscopy can also be used to quantify the concentration of DNA in a sample. DNA absorbs light at 260 nm, so the concentration of DNA in a sample can be determined by measuring the absorbance of light at this wavelength. This technique is commonly used in molecular biology research.
4. **Determination of trace metal concentrations in environmental samples using AAS:** AAS is a highly sensitive technique that can be used to measure trace concentrations of metals in environmental samples. The sample is atomized and the free atoms absorb light at a specific wavelength, which allows the concentration of the metal to be determined. This technique is commonly used in environmental monitoring and quality control.

Overall, analytical spectrophotometry is a powerful and versatile technique that can be used to quantify and identify a wide range of compounds in various types of samples. It has numerous applications in research, industry, and environmental monitoring.

1. **Ultraviolet-visible (UV-Vis) spectroscopy: Properties of light and its interaction with matter**

- Ultraviolet-visible (UV-Vis) spectroscopy is a technique used to measure the absorption of ultraviolet and visible light by a sample. To understand how UV-Vis spectroscopy works, it is important to first understand the properties of light and how it interacts with matter.
- Light is a form of electromagnetic radiation, which travels in waves and is characterized by its wavelength and frequency. The wavelength of light determines its color, with shorter wavelengths appearing blue and violet, and longer wavelengths appearing red and orange. The frequency of light is inversely proportional to its wavelength, meaning that shorter wavelengths have higher frequencies.
- When light passes through a sample, it can be absorbed, transmitted, or reflected. The amount of light that is absorbed or transmitted depends on the properties of the sample, including its chemical composition and structure.
- In UV-Vis spectroscopy, a sample is exposed to light in the UV-Vis range, typically between 200-800 nm. The sample absorbs some of the light, which

reduces the intensity of the light that is transmitted through the sample. The amount of light absorbed by the sample at a specific wavelength can be quantified by measuring the absorbance of the light using a spectrophotometer.

- The amount of light absorbed by a sample at a specific wavelength is proportional to the concentration of the absorbing species in the sample, as well as the path length of the sample and the molar absorptivity of the species. This relationship is described by the Beer-Lambert law:
- $A = \varepsilon lc$
- where A is the absorbance of the light, ε is the molar absorptivity of the absorbing species, l is the path length of the sample, and c is the concentration of the species.
- UV-Vis spectroscopy is commonly used to measure the concentration of chromophores, which are molecules that absorb light in the UV-Vis range. Chromophores include molecules such as proteins, nucleic acids, and pigments. By measuring the absorbance of light at a specific wavelength that corresponds to the chromophore, the concentration of the chromophore in a sample can be determined.

**2. Infrared (IR) spectroscopy: Properties of light and its interaction with matter**

- Infrared (IR) spectroscopy is a technique used to measure the absorption, transmission, and reflection of infrared radiation by a sample. To understand how IR spectroscopy works, it is important to first understand the properties of light and how it interacts with matter.
- Infrared radiation is a type of electromagnetic radiation with longer wavelengths than visible light. It is characterized by its frequency, which is inversely proportional to its wavelength. IR radiation can be divided into three regions: near-IR (NIR), mid-IR (MIR), and far-IR (FIR).
- When IR radiation passes through a sample, it can be absorbed, transmitted, or reflected. The amount of IR radiation that is absorbed or transmitted depends on the properties of the sample, including its chemical composition and structure.

- In IR spectroscopy, a sample is exposed to IR radiation, typically in the mid-IR range (4000-400 cm-1). The sample absorbs some of the IR radiation, which reduces the intensity of the radiation that is transmitted through the sample. The amount of IR radiation absorbed by the sample at a specific wavelength can be quantified by measuring the transmittance or absorbance of the radiation using a spectrophotometer.
- The amount of IR radiation absorbed by a sample at a specific wavelength is proportional to the concentration of the absorbing species in the sample, as well as the path length of the sample and the molar absorptivity of the species. This relationship is described by the Beer-Lambert law:
- $A = \log(Io/I) = \varepsilon lc$
- where A is the absorbance of the IR radiation, Io is the intensity of the incident radiation, I is the intensity of the transmitted radiation, ε is the molar absorptivity of the absorbing species, l is the path length of the sample, and c is the concentration of the species.
- IR spectroscopy is commonly used to identify and characterize the chemical composition of a sample, including functional groups and chemical bonds. Different functional groups and chemical bonds absorb IR radiation at specific frequencies, which produces a unique fingerprint that can be used for identification. By comparing the IR spectrum of an unknown sample to a database of known spectra, the sample can be identified.

### 3. Fluorescence spectroscopy: Properties of light and its interaction with matter

- Fluorescence spectroscopy is a technique used to measure the absorption and emission of light by a sample. To understand how fluorescence spectroscopy works, it is important to first understand the properties of light and how it interacts with matter.
- Light is a form of electromagnetic radiation, which travels in waves and is characterized by its wavelength and frequency. The wavelength of light determines its color, with shorter wavelengths appearing blue and violet, and longer wavelengths appearing red and orange. The frequency of light is

inversely proportional to its wavelength, meaning that shorter wavelengths have higher frequencies.

- When light passes through a sample, it can be absorbed, transmitted, or reflected. The amount of light that is absorbed or transmitted depends on the properties of the sample, including its chemical composition and structure.
- In fluorescence spectroscopy, a sample is exposed to the light of a specific wavelength, which causes the sample to absorb some of the light. This absorbed energy promotes an electron to a higher energy state. The electron then relaxes back to its ground state, emitting light of a longer wavelength than the excitation light. This emitted light is called fluorescence.
- The amount of fluorescence emitted by the sample depends on the concentration and properties of the fluorescent molecules in the sample, as well as the intensity of the excitation light and the detection sensitivity of the instrument. The fluorescence emission spectrum can be measured using a spectrofluorometer.
- Fluorescence spectroscopy is commonly used in biochemistry and biophysics to study the properties and behavior of fluorescent molecules, such as proteins and nucleic acids. Fluorescent molecules can be used as probes to study the interaction of biological molecules with each other and with their environment. By measuring the fluorescence emission of a sample, information about the sample's structure, conformation, and binding interactions can be obtained. Fluorescence spectroscopy can also be used to quantitatively analyse fluorescent molecules, such as in immunoassays or DNA sequencing.

**4. Atomic absorption spectroscopy (AAS): Properties of light and its interaction with matter**

- Atomic absorption spectroscopy (AAS) is a technique used to measure the concentration of atoms in a sample by measuring the absorption of light by the atoms. To understand how AAS works, it is important to first understand the properties of light and how it interacts with matter.
- Light is a form of electromagnetic radiation, which travels in waves and is characterized by its wavelength and frequency. The wavelength of light

determines its color, with shorter wavelengths appearing blue and violet, and longer wavelengths appearing red and orange. The frequency of light is inversely proportional to its wavelength, meaning that shorter wavelengths have higher frequencies.

- When light passes through a sample, it can be absorbed, transmitted, or reflected. The amount of light that is absorbed or transmitted depends on the properties of the sample, including its chemical composition and structure.
- In AAS, a sample is vaporized and atomized using a flame or plasma. The atoms in the sample are then excited by exposing them to light of a specific wavelength, which causes the atoms to absorb some of the light. The amount of light absorbed by the atoms is proportional to their concentration in the sample. The amount of light absorbed by the sample can be quantified by measuring the decrease in the intensity of the incident light using a spectrophotometer.
- The wavelength of the light used in AAS is specific to the element being measured, as different elements absorb light at different wavelengths. By measuring the absorption of light at the characteristic wavelength for a particular element, the concentration of that element in the sample can be determined.
- AAS is commonly used in analytical chemistry to measure the concentration of metal ions in a sample. It is particularly useful for analyzing trace amounts of metals in environmental and biological samples. AAS can also be used to determine the purity of chemicals and to monitor industrial processes.

## Analytical Spectrophotometry: Relation Between Frequency

- Analytical spectrophotometry involves the measurement of the interaction between light and matter, with the goal of determining the chemical composition, structure, or concentration of the sample being analyzed. The frequency of light plays a crucial role in this process, as it determines the energy of the light and how it interacts with the sample.
- The frequency of light is related to its wavelength by the formula $c = \lambda\nu$, where c is the speed of light, $\lambda$ is the wavelength, and $\nu$ is the frequency. This means that light with a shorter wavelength will have a higher frequency and vice versa.

- When light passes through a sample, it can be absorbed, transmitted, or reflected. The amount of light that is absorbed or transmitted by the sample depends on its chemical composition and structure, as well as the frequency of the incident light.
- In fluorescence spectroscopy, a sample is excited by light of a specific frequency, causing it to emit light of a different frequency. The frequency of the emitted light is lower than that of the incident light, due to the energy lost during relaxation of the excited state. The relationship between the frequency of the incident light and the emission of the sample is described by the Stokes shift, which is the difference between the excitation and emission wavelengths.
- In atomic absorption spectroscopy, a sample is vaporized and atomized, and the atoms in the sample are excited by light of a specific frequency. The frequency of the incident light is specific to the element being measured, as different elements absorb light at different frequencies. The amount of light absorbed by the atoms is proportional to their concentration in the sample.
- In summary, the frequency of light plays a critical role in analytical spectrophotometry, as it determines the energy of the incident light and how it interacts with the sample. The relationship between the frequency of the incident light and the absorption, emission, or excitation of the sample depends on the specific spectroscopic technique being used.

## Analytical Spectrophotometry: Relation Between Frequency

- In analytical spectrophotometry, the frequency of light plays a crucial role in determining the chemical composition of a sample. The frequency of light is related to the energy of the light wave, which in turn affects how the light interacts with matter.

- The relationship between frequency and energy is described by the equation E = hν, where E is the energy of the light, h is Planck's constant (6.626 × 10^-34 joule-seconds), and ν is the frequency of the light. This equation shows that the energy of the light is directly proportional to its frequency. In other words, the higher the frequency of the light, the greater the energy of the light wave.

- This relationship between frequency and energy is important in many analytical techniques. For example, in UV-Vis spectroscopy, a sample is exposed to light in the UV and visible regions of the electromagnetic spectrum. The light interacts with the sample, and the amount of light absorbed or transmitted is measured. The amount of light absorbed or transmitted depends on the energy of the light, which is determined by its frequency.

- Infrared (IR) spectroscopy is another analytical technique that relies on the frequency of light. In this technique, a sample is exposed to infrared radiation, which causes molecular vibrations within the sample. The frequency of the infrared radiation is chosen to match the natural frequency of the molecular vibrations in the sample. The amount of light absorbed or transmitted by the sample depends on the strength of these vibrations, which in turn depends on the chemical composition of the sample.

- The relationship between frequency and wavelength is also important in analytical spectrophotometry. The wavelength of light is related to its frequency by the equation $\lambda = c/\nu$, where $\lambda$ is the wavelength, c is the velocity of light, and $\nu$ is the frequency. This means that light with a higher frequency (and shorter wavelength) has a shorter wavelength than light with a lower frequency (and longer wavelength). The relationship between frequency and wavelength is used to determine the range of frequencies or wavelengths that will interact with a given sample in a particular analytical technique.

## Analytical Spectrophotometry: Relation Between Wave Number

- In analytical spectrophotometry, wave number ($\tilde{\nu}$) is a unit of measurement used to describe the frequency of light. The wave number is defined as the reciprocal of the wavelength of the light, multiplied by a factor of 10^7 to convert from meters to centimeters. The relationship between wave number, frequency, and wavelength is given by the equation:
- $\tilde{\nu} = 1/\lambda \times 10^7 = \nu/c \times 10^7$
- where $\tilde{\nu}$ is the wave number, $\lambda$ is the wavelength, $\nu$ is the frequency, and c is the velocity of light.

- The wave number is particularly useful in infrared (IR) spectroscopy, where it is commonly used to describe the frequency of molecular vibrations. In IR spectroscopy, a sample is exposed to infrared radiation, which causes molecular vibrations within the sample. The frequency of the infrared radiation is matched to the natural frequency of the molecular vibrations in the sample. The amount of light absorbed or transmitted by the sample depends on the strength of these vibrations, which in turn depends on the chemical composition of the sample.
- The wave number is a convenient way to describe the frequency of these molecular vibrations because it is directly proportional to the energy of the vibration. Stronger molecular vibrations have higher wave numbers, while weaker vibrations have lower wave numbers. The wave number is also used to identify specific functional groups within a molecule based on their characteristic vibrations.
- In summary, wave number is a unit of measurement used to describe the frequency of light in analytical spectrophotometry. It is particularly useful in IR spectroscopy for describing the frequency of molecular vibrations. The wave number is directly proportional to the energy of the vibration and is used to identify specific functional groups within a molecule based on their characteristic vibrations.

## Lambert-Beer's law

- Lambert-Beer's law (also known as Beer-Lambert law or Beer's law) describes the relationship between the concentration of a solute in a solution and the amount of light absorbed or transmitted by the solution. The absorbance (A) of a solution is directly proportional to the solute concentration (c) and the path length (l) of the solution, and inversely proportional to the amount of light (I) transmitted through the solution, according to the law. The law is stated mathematically as:

- $A = \varepsilon cl$

- where A is the solution's absorbance, is the molar absorptivity or extinction coefficient (a constant that depends on the chemical identity of the solute and the wavelength of light), c is the solute concentration in moles per litre (M), and l is the solution's path length in centimeters.

- The law covers all electromagnetic radiation, including visible, ultraviolet, and infrared light. It is often used in analytical chemistry to calculate the concentration of a solute in a solution based on the solution's absorbance of light at a specific wavelength.

- For example, let's consider a solution of a colored compound that absorbs light at a wavelength of 500 nm. A spectrophotometer is used to measure the absorbance of the solution, which is found to be 0.5. The path length of the solution is 1 cm. The molar absorptivity of the compound at 500 nm is known to be 1000 M^-1 cm^-1. Using Beer's law, we can calculate the concentration of the compound in the solution:

- A = εcl

- 0.5 = (1000 M^-1 cm^-1) x (c) x (1 cm)

- c = 0.0005 M

- Therefore, the concentration of the colored compound in the solution is 0.0005 M.

## Limitations of Lambert-Beer's law

While Lambert-Beer's law is a useful principle for quantitative analysis, it has some limitations that should be considered. Here are some of the limitations:

1. **Valid only within a certain concentration range:** Lambert-Beer's law is valid only within a certain concentration range, typically up to about 1 absorbance unit (AU). The law may not hold properly at higher concentrations due to interactions between the solute molecules, leading to deviations from the linear relationship between absorbance and concentration.

2. **Valid only for monochromatic light:** The law assumes that the light source is monochromatic (i.e., of a single wavelength). In practice, most light sources are polychromatic (i.e., consisting of multiple wavelengths), so the law may not be accurate for all wavelengths.
3. **Valid only for dilute solutions:** The law assumes that the solution is dilute (i.e., the solute molecules are well separated from each other), so that the absorption of each molecule is independent of the others. The law may not hold true for concentrated solutions or with strong intermolecular interactions.
4. **Assumes homogeneity of the sample**: The law assumes that the sample is homogeneous, with no spatial variations in concentration or other physical properties. For samples with spatial variations, the absorbance may vary with position and the law may not be applicable.
5. **Sensitivity to experimental conditions:** Lambert-Beer's law is sensitive to experimental conditions, such as temperature, pH, and solvent composition. Any changes in these conditions can affect the molar absorptivity and the accuracy of the law.

Overall, while Lambert-Beer's law is a useful principle for spectrophotometric analysis, it is important to be aware of its limitations and to ensure that the experimental conditions are appropriate for the application.

## Single beam spectrophotometer

A single-beam spectrophotometer is a type of spectrophotometer that is widely used in analytical chemistry to measure the absorbance of a sample. It measures the absorbance of light at a particular wavelength, which can be used to determine the concentration of a solute in a solution. Here's an explanation of how a single-beam spectrophotometer works:

1. **Light Source:** The spectrophotometer uses a light source that emits a beam of light, typically a tungsten filament lamp, a deuterium lamp, or a light-emitting diode (LED).
2. **Monochromator:** The beam of light is then passed through a monochromator, which is a device that separates the light into its component wavelengths. This is typically done using a diffraction grating or a prism.

3. **Sample Holder:** The monochromatic beam of light is then directed through a sample holder containing the sample. The sample holder is typically a cuvette made of quartz or glass with a path length of 1 cm.
4. **Detector:** After passing through the sample, the light is detected by a photosensitive detector, such as a photodiode or a photomultiplier tube. The detector measures the intensity of the light, which is typically converted to an electrical signal that can be displayed on a computer or a chart recorder.
5. **Calibration:** Before analyzing a sample, the spectrophotometer is calibrated using a blank solution, which is a solvent that does not contain the solute of interest. The blank solution is used to set the instrument to zero absorbance at the selected wavelength.
6. **Absorbance Measurement:** Once the instrument is calibrated, the sample is placed in the sample holder and the absorbance is measured. The absorbance is calculated as the negative logarithm of the ratio of the intensity of the light before and after passing through the sample, according to the Lambert-Beer's law equation $A = \log(I_o/I)$, where A is the absorbance, $I_o$ is the intensity of the incident light, and I is the intensity of the transmitted light.
7. **Calculation of Concentration:** The absorbance of the sample is then used to calculate the concentration of the solute using the Lambert-Beer's law equation $C = A/\varepsilon l$, where C is the concentration, $\varepsilon$ is the molar absorptivity, and l is the path length of the sample holder.

In summary, a single-beam spectrophotometer is a relatively simple device that uses a light source, a monochromator, a sample holder, and a detector to measure the absorbance of a sample. It is widely used in analytical chemistry for quantitative analysis of various solutes in solution.

## Double Beam Spectrophotometer

A double-beam spectrophotometer is a type of spectrophotometer that is commonly used in analytical chemistry for quantitative analysis of various solutes in solution. It is similar to a single-beam spectrophotometer but has two separate beams of light used to simultaneously measure the sample and the reference. Here's a detailed explanation of how a double beam spectrophotometer works:

1. **Light Source:** The spectrophotometer uses a light source that emits a beam of light, typically a tungsten filament lamp, a deuterium lamp, or a light-emitting diode (LED).
2. **Monochromator:** The beam of light is then passed through a monochromator, which is a device that separates the light into its component wavelengths. This is typically done using a diffraction grating or a prism.
3. **Beam Splitter:** The monochromatic beam of light is then directed through a beam splitter, which divides the light into two separate beams: the sample beam and the reference beam.
4. **Sample Holder:** The sample beam is directed through a sample holder containing the sample. The sample holder is typically a cuvette made of quartz or glass with a path length of 1 cm.
5. **Detector:** After passing through the sample, the sample beam is detected by a photosensitive detector, such as a photodiode or a photomultiplier tube. The detector measures the intensity of the light, which is typically converted to an electrical signal that can be displayed on a computer or a chart recorder.
6. **Reference Holder:** The reference beam is directed through a reference holder containing a blank solution, which is a solvent that does not contain the solute of interest. The reference holder is also typically a cuvette made of quartz or glass that has a path length of 1 cm.
7. **Reference Detector:** After passing through the reference holder, the reference beam is detected by another photosensitive detector, which measures the intensity of the light.
8. **Calibration:** The spectrophotometer is calibrated using the blank solution in the reference holder before analysing a sample. The blank solution is used to set the instrument to zero absorbance at the selected wavelength.
9. **Absorbance Measurement**: Once the instrument is calibrated, the sample is placed in the sample holder and the absorbance of the sample beam and the reference beam are simultaneously measured. The absorbance is calculated as the negative logarithm of the ratio of the intensity of the sample beam and the intensity of the reference beam, according to the Lambert-Beer's law equation $A = \log(I_o/I)$, where A is the absorbance, $I_o$ is the intensity of the incident light, and I is the intensity of the transmitted light.

10. **Calculation of Concentration:** The absorbance of the sample is then used to calculate the concentration of the solute using the Lambert-Beer's law equation $C = A/\varepsilon l$, where C is the concentration, ε is the molar absorptivity, and l is the path length of the sample holder.

In summary, a double-beam spectrophotometer is a more advanced type of spectrophotometer that uses two separate beams of light to simultaneously measure the sample and the reference, improving the measurement's accuracy and precision. It is widely used in analytical chemistry for quantitative analysis of various solutes in solution.

## Difference Between Single and Double Beam Spectrophotometers

Single-beam and double-beam spectrophotometers are both commonly used in analytical chemistry for quantitative analysis of various solutes in solution. Still, they differ in how they measure a sample's absorbance. Here are some key differences between the two:

1. **Optical Design:** In a single-beam spectrophotometer, there is only one path for light, which passes through the sample and then to the detector. In contrast, a double-beam spectrophotometer splits the beam of light into two separate paths: one for the sample and the other for a blank/reference solution. This allows for simultaneous measurement of the sample and the reference, which can improve accuracy and precision.
2. **Stability:** A single-beam spectrophotometer is typically less stable than a double-beam spectrophotometer. This is because changes in the intensity of the light source or the detector can affect the measurement's accuracy since no reference measurement is being taken. In contrast, a double-beam spectrophotometer takes a reference measurement, which can help to correct for changes in the intensity of the light source or detector.
3. **Calibration:** A single-beam spectrophotometer requires frequent calibration since the light source and detector can drift over time, leading to inaccurate measurements. In contrast, a double-beam spectrophotometer is typically calibrated once daily since the reference measurement can help correct any drift in the light source or detector.

4. **Sample Throughput:** A single beam spectrophotometer can measure only one sample at a time, which can limit the speed of analysis. In contrast, a double-beam spectrophotometer can measure two samples simultaneously, increasing the analysis speed.
5. **Cost:** Double-beam spectrophotometers are typically more expensive than single beam spectrophotometers due to their more complex optical design and additional components.

In summary, while both single-beam and double-beam spectrophotometers are helpful for the quantitative analysis of solutes in solution, double-beam spectrophotometers offer some advantages over single-beam spectrophotometers in terms of stability, speed of analysis, and accuracy but come at a higher cost.

## Lasers and Lamps as Sources of Light in Spectrophotometry

- Lasers and lamps are two common sources of light used in spectrophotometry, each with its own advantages and limitations.
- Laser sources provide intense, monochromatic light with a narrow wavelength range, making them useful for various analytical applications, such as fluorescence and Raman spectroscopy. They can also provide high power density, which can benefit certain applications requiring high sensitivity or precision.
- However, laser sources can be expensive and may require specialized knowledge to operate and maintain. Additionally, they typically provide a limited range of wavelengths, which can make them less versatile than broad-spectrum light sources like lamps.
- Lamps, on the other hand, provide a broad spectrum of light with a range of wavelengths, which makes them useful for a wide range of applications in spectrophotometry, such as UV-Vis and IR spectroscopy. They are also relatively inexpensive and easy to operate.
- However, lamps may not provide the same level of intensity as laser sources, which can limit their sensitivity and precision in some applications. They may also require frequent replacement, particularly if they are used frequently or at high intensity, which can increase their cost over time.

- In summary, both lasers and lamps have their own advantages and limitations as light sources in spectrophotometry, and the choice of the source depends on the specific application and the analysis requirements.

## Monochromators in Spectrophotometry

- Monochromators are an important component of many spectrophotometers, particularly those used for UV-Vis and fluorescence spectroscopy. A monochromator is a device that selectively filters and isolates a narrow range of wavelengths from a broader light spectrum. It then excites a sample or measures its absorption or emission properties.
- A typical monochromator is made up of three major parts: an entry slit, a diffraction grating or prism, and an exit slit. Only a narrow beam of light enters the monochromator through the entrance slit, which is then dispersed into its constituent wavelengths by the diffraction grating or prism. After that, the exit slit selects only a restricted range of wavelengths to get through to the sample or detector.
- Monochromators can be manually or electronically adjusted to select a specific wavelength range for measurement, which allows for greater flexibility and precision in spectrophotometric analysis. Some advanced spectrophotometers also have multiple or tunable monochromators, which can provide even greater flexibility in selecting specific wavelength ranges.
- Overall, monochromators play a critical role in spectrophotometric analysis by allowing for precise control and selection of the wavelengths used for measurement, which is essential for the accurate and reliable analysis of samples.

## Detectors in Spectrophotometry

In spectrophotometry, detectors are used to measure the intensity of light transmitted or absorbed by a sample at different wavelengths. The choice of the detector depends on the wavelength range of interest and the sensitivity required for the application.

Some commonly used detectors in spectrophotometry include:

1. **Photomultiplier tube (PMT):** This is a high-sensitivity detector that can detect very low levels of light. PMTs are often used in fluorescence spectroscopy because of their high sensitivity to low-level light emissions.
2. **Charge-coupled device (CCD):** This detector is commonly used in UV-Vis spectrophotometers. It consists of an array of photodiodes that can be read out electronically. CCDs have good sensitivity, and their signal-to-noise ratio is also high.
3. **Photodiode array (PDA):** This detector is similar to the CCD but contains multiple photodiodes arranged in a linear array. PDAs are often used in high-resolution UV-Vis spectrophotometers.
4. **Integrating sphere:** This detector measures diffuse reflectance and transmittance. It comprises a hollow sphere with a diffuse reflective inner surface and a small aperture for sample entry.
5. **Infrared detectors:** Infrared detectors are used in infrared spectrophotometers. Several types of infrared detectors include thermocouples, Golay cells, and pyroelectric detectors. These detectors are sensitive to the infrared range of the electromagnetic spectrum.

Overall, the choice of the detector in spectrophotometry depends on the specific application and the range of wavelengths of interest.

## Photomultiplier Tubes in Spectrophotometry

- Photomultiplier tubes (PMTs) are widely used in spectrophotometry due to their high sensitivity and low noise characteristics. PMTs work by converting light photons into electrical signals, which are then amplified by a series of dynodes.
- In spectrophotometry, PMTs are often used in applications where the sample emits low light levels, such as in fluorescence or Raman spectroscopy. PMTs are also used in detectors for liquid chromatography and capillary electrophoresis.
- PMTs have several advantages over other detectors in spectrophotometry. They have high sensitivity to low levels of light, a wide dynamic range, and a fast response time. They are also less affected by ambient light and have low dark currents.

- However, PMTs also have some limitations. They are more expensive than other detectors, can be affected by magnetic fields, and require a high-voltage power supply. They are also sensitive to temperature changes, which can affect their stability and performance.

## Photodiode Array

- A photodiode array (PDA) is a type of detector used in spectrophotometry that contains an array of photodiodes arranged in a linear pattern. PDAs are commonly used in high-resolution UV-Vis spectrophotometers.
- When light enters a PDA, it is dispersed by a diffraction grating or prism, and each wavelength of light falls on a different photodiode. The electrical current generated by each photodiode is proportional to the amount of light that falls on it. The resulting signal from each photodiode is then combined to produce a spectrum of the sample.
- PDAs offer several advantages over other detectors in spectrophotometry. They have a high spectral resolution, which makes them useful for analyzing narrow spectral features. They also have a fast acquisition time and a large dynamic range, which allows them to detect a wide range of light intensities.
- One limitation of PDAs is that they are more expensive than some other detectors, such as photomultiplier tubes or charge-coupled devices. They are also more complex to operate and require specialized electronics for signal processing. However, their high spectral resolution and fast acquisition time make them ideal for applications that require high precision and speed, such as in pharmaceutical or environmental analysis.

## Charged Couple Device

- A charged-coupled device (CCD) is a type of detector used in spectrophotometry that consists of an array of photodiodes that can be read out electronically. Each photodiode in the array is sensitive to a specific wavelength range of light. When light strikes the photodiodes, it creates an electrical charge that is then transferred from one photodiode to the next in a process known as "charge-coupling."

- CCDs are widely used in UV-Vis spectrophotometers because they offer several advantages over other types of detectors. They have high sensitivity, fast response times, and low noise. They also have a high signal-to-noise ratio, making them ideal for low-level light detection.

- CCDs are particularly useful in applications that require detecting weak signals, such as in fluorescence spectroscopy or analysing samples with low concentrations. They can also measure multiple wavelengths simultaneously, making them ideal for spectral scanning.

- One limitation of CCDs is that they are sensitive to temperature fluctuations, which can affect their performance. They also have a smaller dynamic range than other types of detectors, such as photomultiplier tubes (PMTs). However, the high sensitivity and low noise of CCDs make them an excellent choice for many types of spectrophotometry applications.

## FT-IR Spectroscopy

- Fourier Transform Infrared (FT-IR) Spectroscopy is a technique used in analytical chemistry to obtain information about the vibrational modes of molecules. It is based on the interaction of infrared radiation with matter, and the resulting absorption or transmission spectra can be used to identify and quantify the presence of various functional groups and chemical bonds in a sample.
- In FT-IR spectroscopy, a beam of infrared light is directed through the sample and the resulting transmitted or absorbed light is measured. The signal is then analyzed using a Fourier Transform algorithm to generate a spectrum that shows the intensity of the absorbed or transmitted light as a function of the frequency or wavelength of the radiation. The resulting spectra are highly specific to the sample's chemical composition and can be used for qualitative and quantitative analysis.
- FT-IR spectroscopy is widely used in a range of applications, including pharmaceuticals, polymers, food, and environmental analysis. It can be used to analyze solids, liquids, and gases, and is particularly useful for analysing complex mixtures.

- One major advantage of FT-IR spectroscopy is its high sensitivity and selectivity. It can detect very small amounts of analytes and provide detailed information about their chemical structure. It is also a non-destructive technique, meaning samples can be analyzed without alteration or destruction.
- However, FT-IR spectroscopy does have some limitations. It is often limited in its ability to provide information about the spatial distribution of analytes, and it may not be suitable for analysing highly absorbing or scattering samples. Additionally, the interpretation of spectra can be complex and requires expertise in the field of vibrational spectroscopy.

## Fourier Analysis in FT-IR

- In FT-IR spectroscopy, Fourier analysis is used to transform the raw data from a time-domain signal (the interferogram) into a frequency-domain spectrum. This is done using a mathematical technique called a Fourier transform, which converts the interferogram data from the time domain into the frequency domain.
- The interferogram is obtained by measuring the interference pattern of a beam of infrared light that has passed through a sample and a reference. This interference pattern contains information about the absorption of the sample at different frequencies. By performing a Fourier transform on the interferogram, this information can be extracted and presented in the form of a frequency-domain spectrum.
- The resulting FT-IR spectrum shows the intensity of the absorbed or transmitted light as a function of the frequency or wavelength of the radiation. The peaks in the spectrum correspond to the different vibrational modes of the molecular bonds in the sample, and can be used to identify and quantify the presence of various functional groups and chemical bonds in the sample.
- In FT-IR spectroscopy, Fourier analysis is an essential part of data processing and interpretation, allowing for the extraction of valuable information from the raw interferogram data.

## Interferometry in FT-IR

- Interferometry is a technique used in FT-IR spectroscopy to obtain a highly precise and accurate measurement of the transmitted or reflected light intensity from a sample. In FT-IR spectroscopy, interferometry is used to measure the interference pattern between a sample beam and a reference beam of infrared light.
- The interferometer is the key component of an FT-IR spectrometer, and it consists of several mirrors that split the incoming infrared light into two beams. One beam is directed to the sample, while the other is directed to a reference mirror. After reflecting off the sample and the reference mirror, the beams are recombined at a detector, which measures the resulting interference pattern.
- The interference pattern is known as an interferogram, containing information about the sample's absorbance or transmittance at different frequencies. The interferogram is then processed using a Fourier transform to generate a spectrum that shows the intensity of the absorbed or transmitted light as a function of the frequency or wavelength of the radiation.
- Interferometry in FT-IR spectroscopy has several advantages. It provides highly accurate and precise measurements of the intensity of the transmitted or reflected light, and it is capable of measuring the entire spectrum of the sample in a single measurement. This allows for rapid and efficient analysis of complex samples.
- However, interferometry also has some limitations. It requires a highly stable and accurate interferometer, which can be expensive and difficult to maintain. Additionally, interferometry is sensitive to environmental factors such as temperature and vibration, which can affect the accuracy and precision of the measurements.

## Noise and Control in FT-IR

- Noise is a common issue in FT-IR spectroscopy, and it can significantly affect the accuracy and precision of the measurements. In FT-IR spectroscopy, there are several sources of noise, including electronic noise, detector noise, and background noise.

- Electronic noise can be reduced by using high-quality electronic components, such as low-noise amplifiers and analog-to-digital converters. Detector noise can be reduced using high-quality detectors with low noise levels, such as liquid-nitrogen-cooled mercury cadmium telluride (MCT) detectors. Background noise can be reduced by using appropriate background subtraction techniques, such as the use of a reference sample or a blank sample.
- In addition to these hardware-based noise reduction techniques, there are also several software-based noise reduction techniques that can be used in FT-IR spectroscopy. These include averaging multiple scans, smoothing the spectrum using digital filters, and using advanced signal processing techniques like wavelet transforms.
- To control noise in FT-IR spectroscopy, it is important to design and optimize the experimental setup carefully, and to use appropriate noise reduction techniques during data processing. This may include optimizing the sample preparation and handling, selecting appropriate instrumental settings, and using appropriate data processing techniques.
- Overall, controlling noise in FT-IR spectroscopy is essential for obtaining accurate and precise measurements, and it requires careful attention to the experimental design and data processing techniques.

## Classical Analytical methods: Acid-Base

- The acid-base titration method is a traditional analytical method for determining the concentration of an unknown acid or base solution by reacting it with a known concentration of acid or base solution. The acid-base process is a neutralization reaction that creates salt and water. When the acid and base are totally neutralized and the solution is neither acidic nor basic, the reaction has achieved its equivalence point.
- The acid-base titration method requires several steps. First, a known volume of the unknown solution is placed in a flask or beaker. A suitable indicator is added to the solution to help determine when the equivalence point is reached. Then, a known concentration of the titrant (either an acid or a base) is slowly added to the solution while stirring until the endpoint is reached, as indicated by the change in color of the indicator.

- The endpoint of the reaction is the point at which the acid and base are completely neutralized, and the solution is neither acidic nor basic. At this point, the amount of titrant added equals the amount of the unknown solution present. The concentration of the unknown solution can then be calculated using the stoichiometry of the neutralization reaction and the volume and concentration of the titrant solution.
- An acid-base titration is a widely used analytical method due to its simplicity, speed, and accuracy. It can be used to determine the concentration of a wide range of acids and bases and is particularly useful for determining the concentration of weak acids or bases, which are difficult to measure using other analytical methods.

## Complexometric and Redox Titrations

- Complexometric titration is a classical analytical method used to determine the concentration of metal ions in a solution by reacting them with a known amount of a chelating agent (complexing agent) that forms a complex with the metal ions. The chelating agent typically has multiple donor atoms that form coordination bonds with the metal ion. The formation of the metal-chelate complex is indicated by a color change or other observable property.
- The endpoint of the complexometric titration is reached when all of the metal ions have reacted with the chelating agent and are present in the form of the metal-chelate complex. The concentration of the metal ion can then be calculated based on the stoichiometry of the reaction and the amount of chelating agent added.
- Redox titration is another classical analytical method used to determine the concentration of a substance in a solution by reacting it with a known amount of an oxidizing or reducing agent. Redox titrations are typically used to determine the concentration of substances that are capable of undergoing oxidation or reduction reactions. The endpoint of the redox titration is reached when all of the substance being analyzed has reacted with the oxidizing or reducing agent and the stoichiometric equivalence point has been reached.
- Redox titrations can be carried out using a variety of oxidizing or reducing agents, depending on the substance being analyzed. Common oxidizing agents

used in redox titrations include potassium permanganate, potassium dichromate, and iodine. Common reducing agents used in redox titrations include sodium thiosulfate, ascorbic acid, and ferrous ammonium sulphate.

- Overall, complexometric and redox titrations are widely used classical analytical methods due to their simplicity, speed, and accuracy. They can be used to determine the concentration of a wide range of substances and are particularly useful for analyzing metal ions and substances capable of undergoing oxidation or reduction reactions.

## Gravimetric Analysis

- Gravimetric analysis is a classical analytical method used to determine the amount of a substance present in a sample by measuring its mass. This method involves isolating the substance of interest in a pure form, typically by precipitation or volatilization, and then weighing the resulting solid or gas.
- In gravimetric analysis, a known sample volume is taken, and the substance of interest is separated from the other components of the sample. The separated substance is then collected and weighed, and the mass of the substance is used to calculate its concentration in the original sample.
- Gravimetric analysis requires careful attention to experimental details to ensure accurate and precise results. The sample must be carefully prepared and weighed, and the conditions of the separation process must be controlled to ensure complete isolation of the substance of interest. In addition, the purity of the separated substance must be confirmed through additional tests, such as melting point determination or chemical analysis.
- Gravimetric analysis is commonly used in the analysis of metals, particularly in the determination of the composition of alloys. It is also used in the analysis of inorganic compounds, such as salts and minerals. While gravimetric analysis is a time-consuming and labor-intensive method, it is a highly accurate and precise method for the determination of the amount of a substance present in a sample.

## References


1. "Quantitative Chemical Analysis" by Daniel C. Harris
2. "Principles of Instrumental Analysis" by Douglas A. Skoog, F. James Holler, and Stanley R. Crouch
3. "Fundamentals of Analytical Chemistry" by Douglas A. Skoog, Donald M. West, F. James Holler, and Stanley R. Crouch
4. "Analytical Chemistry: A Modern Approach to Analytical Science" by David Harvey
5. "Instrumental Methods of Analysis" by Hobart H. Willard, Lynne L. Merritt, Jr., John A. Dean, and F. James Holler


## About The Authors


**Dr. Yasmeen JUNEJO**, is an Assistant Professor of Physiology and Biochemistry at Cholistan University of Veterinary and Animal Sciences (CUVAS), Bahawalpur, Pakistan. He holds a PhD in Analytical Chemistry from Sindh University, Jamshoro, Pakistan and Postdoc from Gaziantep University, Turkey. His main areas of interest are analytical chemistry, nanomedicine, nanostructures, nanobiotechnology and their applications. She is an editor of international journals and also written many books and book chapters as well.

E-mail:yasmeen@cuvas.edu.pk, ORCID:0000 0002 3720 3214

**Prof. Dr. Mehmet ÖZASLAN** received his PhD in 1995 Institute of Natural Sciences at Cukurova University, Turkiye. He is a Professor in Molecular Biology and Genetics. His research interests are included Cancer Genetics, Molecular Virology, Molecular Genetics, Microbiology, and Genetic mutations etc. He has published more than 200 research articles in national and international well reputed journals. He also has written many book chapters as well as edited books. He is an editor and editor in chief of many well reputed national and international journals.
E-mail: ozaslanmd@gantep.edu.tr , ORCID: 0000 0001 9380 4902



To Cite This Chapter :
Junejo, Y. & Ozaslan, M. (2023). **Basics to advanced analytical chemistry**. In Y. Junejo (Ed.), *Chemistry for the Life* (pp. 247-290). ISRES Publishing.